



FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmaglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

3000TA, Process Trace Oxygen Analyzer.
NI / I / 2 / BCD / T3 Ta = 50°C

Special Condition of Use:

1. The product shall be installed in compliance with the enclosure, mounting, spacing and segregation requirements of the ultimate application including access only by the use of a tool.

Equipment Ratings:

Non-incendive for use in Class I, Division 2, Groups B, C and D Hazardous (Classified) Locations

Approved for:

Teledyne Analytical Instruments
18830 Chestnut St.
City of Industry, CA 91719



This certifies that the equipment described has been found to comply with the following FM Approval Standards and other documents:

Class 3600	1998
Class 3611	1999
Class 3810	1989


Original Project ID: 3014276

FM Approval Granted: October 3, 2002

Subsequent Revision Reports / Date FM Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Global Technologies LLC



David W. Styrula
Technical Team Manager

10/9/02
Date