

EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer

APPLICATION OF COUNCIL DIRECTIVE : 2014/34/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN 60079-0 : 2012 (IEC 60079-0 : 2011)
EN 60079-11 : 2012 (IEC 60079-11 : 2011)
EN 60079-26 : 2007 (IEC 60079-26 : 2006)

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Portable Oxygen Analyzer 3110

EQUIPMENT CLASS : Electrical Equipment Measurement, Control & Laboratory Use -Industrial

MODEL NUMBER : 3110

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S) PER 2014/34/EU and have been type-approved by ISSeP, Zoning A. Schweitzer, rue de la Platinerie, B-7340 Colfontaine. ISSeP notified body identification number: 0492. Mark: Ex II 1 G. Code: Ex ia II C T3 Ga. EC Type Certificate Number: ISSeP08ATEX036X


SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 7-27-18

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 7-27-18

PLACE: City of Industry, California