

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL : 2014/30/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED: EN61326-1: 2013
EN55011 Class A Group 1

EN61000-4-2, EN61000-4-3, EN61000-4-4,
EN61000-4-5, EN61000-4-6, EN61000-4-8
EN61000-4-11

EN61010-1: 2010 LVD

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Oxygen Analyzer

EQUIPMENT CLASS : Electrical Equipment Measurement,
Control & Laboratory Use- Industrial

MODEL NUMBER : 3000MA

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE DIRECTIVE(S) AND STANDARD(S).

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

DATE: 4-11-17

PLACE: City of Industry, CA