

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL : 94/9/EC
ATEX DIRECTIVE

STANDARDS TO WHICH : EN 60079-0:2006
CONFORMITY IS DECLARED : EN 60079-11:2007
EN 60079-26:2007

MANUFACTURER'S NAME : Teledyne Analytical Instruments

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Oxygen Measuring Transmitter
Type Insta Trans

EQUIPMENT CLASS : Ex II 2 (1) G EEx ia IIC T4 (0°C < T_a < 50°C)

MODEL NUMBER : 3001 Insta Trans

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 94/9/EC and have been type-approved by Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number: 1180. Mark: Ex II 2 (1) G. Code: EEx ia IIC T4 (0°C < T_a < 50°C).

SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 3-14-08

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 3/14/08

PLACE: City of Industry, California