

# EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer

---

APPLICATION OF COUNCIL : 2014/34/EU  
ATEX DIRECTIVE

STANDARDS TO WHICH : EN61326-1: 2013  
EN55011 Class A Group 1  
EN 60079-0:2012+All:2014  
(IBC 60079-0:2011)

CONFORMITY IS DECLARED : EN 60079-11:2012  
(IEC 60079-11:2011)

MANUFACTURER'S NAME : Teledyne Analytical Instruments

MANUFACTURER'S ADDRESS : 16830 Chestnut Street  
City of Industry, CA 91748  
U.S.A.

TYPE OF EQUIPMENT : Oxygen Measuring Transmitter  
Type Insta Trans-XD

EQUIPMENT CLASS : Ex II IG Ex ia IIC T4 Ga  
(0°C ≤ Ta ≤ +50°C)


MODEL NUMBER : Insta Trans-XD


---

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards per 2014/34/EU and have been type-approved by SGS Fimko Oy, Takomotie 8, FI-00380 Helsinki, Finland. SGS Fimko Oy, notified body identification number 0598. Mark: Ex II IG. Code: Ex ia IIC T4 Ga

EC Type Certificate Number: Baseefa12ATEX0114X

IECEX Certificate Number: IECEX BAS 12.0078X

SIGNATURE:   
FULL NAME: Nino Lague  
POSITION: QA Analyst  
Date: 12-13-22  
PLACE: City of Industry, California

SIGNATURE:   
FULL NAME: Angel Alegria  
POSITION: New Products Engineer  
Date: 12-13-22  
PLACE: City of Industry California