

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE : Directive 2014/34/EU
Directive 2014/30/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED : IEC 60825-1, IEC 62321-1:2013, IEC 62321-2:2013, IEC 62321-3-1:2013, IEC 62321-4:2013, IEC62321-5:2013, IEC 62321-6:2015, IEC 61010-1:2010, IEC 60825-1:2014, EN 61326-1:2013, IEC 61000-4-2:2008, IEC 61000-4-3:2010, IEC 61000-4-4:2010, IEC 61000-4-5:2005, IEC 61000-4-6:2008, IEC 61000-4-8:2009, IEC 60079-0:2011, IEC 60079-1:2014, IEC 60079-28:2015, IEC 60079-31:2013, IEC 80079-34:2011

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Laser Gas Analyzer

EQUIPMENT CLASS : Laboratory, Measurement & Process Control

MODEL NUMBER : LGA4000Z

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S)

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 7/14/20

PLACE: City of Industry, California