

# EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer

---

**APPLICATION OF COUNCIL ATEX DIRECTIVE** : 2014/34/EU

**STANDARDS TO WHICH CONFORMITY IS DECLARED** : EN 60079-0:2012+A11:2014  
(IEC 60079-0:2011)  
EN 60079-11:2012  
(IEC 60079-11:2011)

**MANUFACTURER'S NAME** : Teledyne Analytical Instruments

**MANUFACTURER'S ADDRESS** : 16830 Chestnut Street  
City of Industry, CA 91748  
U.S.A.

**TYPE OF EQUIPMENT** : Oxygen Measuring Transmitter  
Type Insta Trans-XD

**EQUIPMENT CLASS** : Ex II 1G Ex ia IIC T4 Ga  
(0°C ≤ Ta ≤ +50°C)

**MODEL NUMBER** : Insta Trans-XD

---

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 2014/34/EU and have been type-approved by SGS Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number:1180. Mark: Ex II 1G. Code: Ex ia IIC T4 Ga  
EC Type Certificate Number: Baseefa12ATEX0114X  
IECEX Certificate Number: IECEX BAS 12.0078X

SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 7-27-18

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 7-27-18

PLACE: City of Industry, California