

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL : 89/336/EEC
73/23/EEC

**STANDARDS TO WHICH
CONFORMITY IS DECLARED** : EN55011-Class A Group I
EN50082-2
EN61010-1

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Oxygen Analyzer

EQUIPMENT CLASS : ISM Class A Group 1

MODEL NUMBER : 3010MB

**I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT
SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S) PER
89/336/EEC AND 73/23/EEC.**

SIGNATURE: Stephen Broy

FULL NAME: Stephen Broy

POSITION: Director of Engineering

DATE: 12-16-13

PLACE: City of Industry, CA