

EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer

APPLICATION OF COUNCIL : 89/336/EEC
ATEX DIRECTIVE : 2014/34/EU

STANDARDS TO WHICH : EN 60079-0:2012
CONFORMITY IS DECLARED : EN 60079-11:2012

MANUFACTURER'S NAME : Teledyne Analytical Instruments



MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Portable Oxygen Analyzer

EQUIPMENT CLASS : ISM Class A Group 1

MODEL NUMBER : 311TC, 311PC & 311TCXL

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 89/336/EEC and 2014/34/EU and have been type-approved by SGS Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number: 1180. Mark: Ex II 2 G. Code: Ex ib IIC T3 Gb. EC Type Certificate Number: BAS02ATEX2303

SIGNATURE:  SIGNATURE: 

FULL NAME: Roger Starlin FULL NAME: Angel Alegria

POSITION: QA Manager POSITION: New Products Manager

Date: 7-27-18 Date: 7-27-18

PLACE: City of Industry, California PLACE: City of Industry, California