

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE : 89/336/EEC (92/31/EEC)
73/23/EEC

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN55011.1:1995 (ISM, Group 1, Class B)
EN55014:1995 (Clause 4.2)
EN50082-2 (Industrial)
AS61000.4.5:1999
AS1099.2.6
IEC-68-2-2
IEC-68-2-3

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Oxygen Analyzer

MODEL NUMBER : 9060

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE DIRECTIVE(S) AND STANDARD(S).

SIGNATURE:  _____

FULL NAME: Stephen Broy

POSITION: Director of Engineering

DATE: 7-3-06

PLACE: City of Industry, CA